

AQUAMEAS INST PVT LTD		DATA SHEET FOR ULTRASONIC TYPE LEVEL MEASURING SYSTEM	DS No	INST/DS/07	
				REV	0
	Date		#####		
Checked by	DSS				
Approved by	KPJ				
Revision details					
Sr No	Particular	Description			
General					
1	Make	International Standard			
2	Service	Water level measurements			
3	Quantity & Range	upto 12 Mtrs			
4	Accuracy of measuring loop	$\pm 0.5\%$ of full scale			
Sensor/transmitter					
1	Output	4-20 mA			
2	Mounting	From top of reservoir, tank			
3	Programming facility with programmer	Required			
4	Display unit	Avalibale integral			
Note :					